



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
LR1405

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	AirSep Corporation
Address	260 Creekside Drive Buffalo, NY 14228
Project Number	LR1405-2R3
Product	Oxygen Generator
Model Number	Topaz (AS013-105, AS013-115, AS013-118, AS013-124), Topaz Ultra (AS123-7, AS123-8, AS123-9, AS123-10)
Ratings	110/120Vac, 60Hz, 5.0A 110/120Vac, 60Hz, 6.0A
Applicable Standards	<ul style="list-style-type: none"> • CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements - Second Edition; Incorporating Update No. 1: October 2008 • UL 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements Including revision October 2008
Manufacturing/Factory Location	AirSep Corporation 260 Creekside Drive Buffalo, NY 14228

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Scott Airdrie

Signature:

Date: February 11, 2019