



QPS Evaluation Services Inc
 Testing, Certification and Field Evaluation Body
 Accredited in Canada, the USA, and Internationally

File
LR1405

CERTIFICATE OF COMPLIANCE
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	AirSep Corporation
Address	260 Creekside Drive Buffalo, New York 14228, USA
Project Number	LR1405-3R1
Product	Oxygen Concentrator
Model Number	AS074-1 Centrox
Ratings	Cord connected, 120Vac, 11A, 60Hz for indoor use only.
Applicable Standards	<ul style="list-style-type: none"> • CAN/CSA-C22.2 No. 60601-1:08 - Medical electrical equipment – Part 1: General requirements for basic safety and essential performance - Second Edition; Incorporates Corrigendum 2: June 2011 • AAMI ES60601-1 - Medical electrical equipment - Part 1: General requirements for basic safety and essential performance - Consolidated reprint C1: 2009; Incorporates Amendment 2: 2010
Factory/Manufacturing Location	AirSep Corporation 260 Creekside Drive Buffalo, New York 14228, USA

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Scott Airdrie

Signature:

Date: May 25, 2018