



QPS Evaluation Services Inc
 Testing, Certification and Field Evaluation Body
 Accredited in Canada, the USA, and Internationally

File
LR1405

CERTIFICATE OF COMPLIANCE
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Airsep Corporation
Address	260 Creekside Drive, Buffalo, NY, 14228, USA
Project Number	LR1405-1R3
Product	Oxygen concentrators
Model Number	ONYX AS016-1, ONYX+ AS017-1 and ONYX ULTRA AS121-1
Ratings	ONYX AS016-1 and ONYX+ AS017-1, rated 120Vac, 60Hz, 5.0A ONYX ULTRA AS121-1, rated 120Vac, 60Hz, 6.0A
Applicable Standards	<ul style="list-style-type: none"> CAN/CSA-C22.2 No. 61010-1-12- Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements - Third Edition; Update No. 1: July 2015 UL 61010-1 - UL Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements - Third Edition; Reprint with revisions through and including July 15, 2015
Factory/Manufacturing Location	CAIRE Inc. 2205 Airport Industrial Drive Ball Ground, GA 30107

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Andrew Saunders

Signature:

Date: February 8, 2021