



SERVICE BULLETIN

PN: 21297437

RELEASE DATE: February 27, 2019

REVISION

DATE(S): June 28, 2018 (rev A), February 27, 2019 (rev B)

MODEL: CAIRE LOX Products

ISSUE: IEC 60601-1-2 Medical Electrical Equipment – Electromagnetic Disturbances 4th Edition Compliance

NOTES: All currently manufactured CAIRE LOX products with electronic liquid level gauges are IEC 60601-1-2 3rd Edition compliant. Please see table below for complete list of current models. Note: products that do not have an electronic liquid level gauge are not required to be compliant with 60601-1-2.

Reservoirs				
Liberator 20	Liberator 30	Liberator 37	Liberator 45	Liberator 60
Helios H36	Helios H46	Helios U36	Helios U46	
Portables				
Spirit 300	Spirit 600	Spirit 1200		
Sprint	Stroller	Hi Flow Stroller		

The above mentioned CAIRE LOX products will be IEC 60601-1-2 4th Edition compliant in anticipation of the 4th edition which will be required in Europe as of Jan 1st, 2019, with the exception of the Helios Reservoir product lines will have already been or will be discontinued by that time (see bulletins 21267961 and 21285171 for details on those announcements).

CONTACT: For technical questions or concerns, contact Technical Service:

USA: 800-482-2473

www.cairemedical.com

Europe: +44 (0) 118 936 7060

www.cairemedical.com

For ordering information or general inquires, contact Customer Service:

USA: 800-482-2473

www.cairemedical.com

Europe: +44 (0) 118 936 7060

www.cairemedical.com

Copyright © 2019 CAIRE Inc. CAIRE Inc. reserves the right to discontinue its products, or change the prices, materials, equipment, quality, descriptions, specifications and/or processes to its products at any time without prior notice and with no further obligation or consequence. All rights not expressly stated herein are reserved by us, as applicable.