



QPS Evaluation Services Inc
 Testing, Certification and Field Evaluation Body
 Accredited in Canada, the USA, and Internationally

File LR1405

CERTIFICATE OF COMPLIANCE
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	AirSep Corporation
Address	260 Creekside Drive Buffalo, NY 14228, USA
Project Number	LR1405-2R5
Product	Oxygen Generator
Model Number and Ratings	<ul style="list-style-type: none"> • Topaz (AS013-105, AS013-115, AS013-118, AS013-124); rated 110/120Vac, 60Hz, 5.0A • Topaz Ultra (AS123-7, AS123-8, AS123-9, AS123-10); rated 110/120Vac, 60Hz, 6.0A
Applicable Standards	<ul style="list-style-type: none"> • CAN/CSA-C22.2 No. 61010-1-04: Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements - Second Edition; Incorporating Update No. 1: October 2008 • UL 61010-1: Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements; Second Edition including revision October 2008 •
Manufacturing/Factory Location	AirSep Corporation 260 Creekside Drive Buffalo, NY 14228, USA

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Simon Hodson

Signature: *Simon Hodson*

Date: December 07, 2020