



**QPS Evaluation Services Inc**  
**Testing, Certification and Field Evaluation Body**  
**Accredited in Canada, the USA, and Internationally**

File  
LR1405

**CERTIFICATE OF COMPLIANCE**  
**(ISO TYPE 3 CERTIFICATION SYSTEM)**

Issued to	Airsep Corporation
Address	260 Creekside Drive, Buffalo, NY, 14228, USA
Project Number	LR1405-1R4
Product	Oxygen concentrators
Model Number	ONYX AS016-1, ONYX+ AS017-1 and ONYX ULTRA AS121-1
Ratings	ONYX AS016-1 and ONYX+ AS017-1, rated 120Vac, 60Hz, 5.0A ONYX ULTRA AS121-1, rated 120Vac, 60Hz, 6.0A
Applicable Standards	<ul style="list-style-type: none"> <li>• <b>CAN/CSA-C22.2 No. 61010-1-12:</b> Electrical equipment for measurement, control, and laboratory use - Part 1: General requirements - Third Edition; Update No. 1: July 2015</li> <li>• <b>UL 61010-1:</b> Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements - Third Edition; Reprint with revisions through and including July 15, 2015</li> </ul>
Factory/Manufacturing Location	CAIRE Inc. 2205 Airport Industrial Drive Ball Ground, GA 30107 USA

**Statement of Compliance:** The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Scott Airdrie

Signature: *Scott Airdrie*

Date: February 28, 2022