

Declaration of Conformity MDD, RoHS 3



Legal Manufacturer / Address	CAIRE Inc. 2200 Airport Industrial Drive Suite 500 Ball Ground, GA 30107 USA		
Manufacturing	CAIRE Inc. 2205 Airport Industrial Drive Ball Ground, GA 30107 USA		
Repair Facilities	CAIRE Inc. 2205 Airport Industrial Drive Ball Ground, GA 30107 USA	I/F, Cro Dev	IRE Medical Technology (Chengdu) Co. Ltd. Building 4, No 670 Haifa Road, Chengdu bss-Strait Science and Technology Industrial velopment Park Wenjiang District, Chengdu huan Province, 611138 China
Distribution & order fulfilment	CAIRE Medical Germany GmbH Arnold-Höveler-Strasse 2, 40764 Langenfeld. Germany		
Notified Body	SGS Belgium NV (1639) Noorderlaan 87 BE-2023 Antwerton Belgium	1, r 750	MED SAS (0459) – Certificate Only. rue Gaston Boissier 015 PARIS ance
Authorized Representative	Medical Product Service GmbH (MPS) Borngasse 20 35619 Braunfels, Germany		Accumed Sagl Viale Serfontana 10 CH-6834 Morbio Inferiore, Switzerland.
Product Families	Oxygen Therapy Flowmeter Stand – Model SureFlow		
MDD Device Classification	Class IIa (MDD Appendix IX, Rule 2)		
Global Medical Device Nomenclature (GMDN) Code	61365 - Oxygen Therapy Flowmeter		
Start of CE Marking	3 April 2019		
Technical File Reference	SUREFLOW-MDD-TF-01		
Effective Date	May 15, 2024		
Expiration Date	Dec 31, 2028		

This declaration applies to all CE marked devices manufactured from date of issuance until it is either superseded by another declaration or withdrawn. I, the undersigned, hereby declare the equipment specified above meets the essential requirements and conforms to the following:

- Council Directive 93/42/EEC, Medical Devices Directive (Annex II, excluding section 4) including Amendment 2007/47/EC, and Article 120 Transitional provisions under MDR (2017/745) as revised by EU regulation 2023/607. GMED SAS Certificate No° 31275 rev. 14, Expiry Date: May 26, 2024, is hereby extended by confirmation letter reference CLNB1639-WW/MC/625801 from SGS.
- EN ISO 13485:2016, Quality Management Systems (Certificate 31276)

These devices also conform to the following Directive(s):

 Council Directive 2011/65/EU, Restriction of hazardous substances; Directive (EU) 2015/863, amending Annex II to 2011/65/EU

Columpalal Cross

Date: May 15, 2024