



SERVICE BULLETIN

PN: 21250054

RELEASE DATE: February 19, 2019

REVISION DATE(S): April 18, 2018 (rev B), February 19, 2019 (rev C)

MODEL: Eclipse 2, Eclipse 3 Oxygen Concentrators

ISSUE: Eclipse 2 End of Service Life and Eclipse 3 Discontinuation/End of Service Life

NOTES: Effective immediately, CAIRE will no longer be offering service options for the Eclipse 2 Oxygen Concentrator. This includes internal service offerings, as well as service components. The Eclipse 2 does share the same compressor (PN: SP5999-SEQ) and ATF assembly (PN: SP20626702) as the current model, the Eclipse 5. Limited Eclipse 2 specific components will be available until inventory has been depleted. Effective August 31, 2020, CAIRE will no longer be offering service options for the Eclipse 3 Oxygen Concentrator. The Eclipse 3 was discontinued in August 2013, and replaced by the current model, the Eclipse 5. Parts and service will be available for units still within the warranty period, and parts and service will continue to be available at Chart's discretion.
The French version Eclipse 3 (PN: 5901-SEQ) was discontinued December 2017, and the End of Service Life will be December 31, 2022

CONTACT: For technical questions or concerns, contact Technical Service:

USA: 800-482-2473

www.cairemedical.com

Europe: +44 (0) 118 936 7060

www.cairemedical.com

For ordering information or general inquiries, contact Customer Service:

USA: 800-482-2473

www.cairemedical.com

Europe: +44 (0) 118 936 7060

www.cairemedical.com

Copyright © 2019 CAIRE Inc. CAIRE Inc. reserves the right to discontinue its products, or change the prices, materials, equipment, quality, descriptions, specifications and/or processes to its products at any time without prior notice and with no further obligation or consequence. All rights not expressly stated herein are reserved by us, as applicable.